



**CITY OF ST. LOUIS
DEPARTMENT OF FINANCE
OFFICE OF THE SUPPLY COMMISSIONER**

CAROL L. SHEPARD
SUPPLY COMMISSIONER

FRANCIS G. SLAY
MAYOR

1200 MARKET ST RM 324
SAINT LOUIS MO 63103
PHONE 314-622-4580
FAX 314-622-4141

**ADDENDUM NO. 1
BID OPENING: NOVEMBER 25, 2014**

TO: ALL BIDDERS
FROM: LYNN CRAWFORD, BUYER
DATE: NOVEMBER 19, 2014
RE: RFQ 33515Q0006, Autopsy Saw

This addendum is issued prior to bid opening to provide for modifications in plans and/or specifications. All bidders must attach this form to the original Bid/Quote Form. This completed, signed form will serve as acknowledgement that revised information has been received and the Bidder's bid is based on the revised specifications.

1. QUESTIONS/ANSWERS:

Question 1: The Mortech Saw on the RFQ is no longer made. Is there another manufacturer you can reference?

Answer: The specifications are autopsy saw that cuts bone by high speed oscillation. Operates on 115 Vac, 60 Hz. Saw shall be CSA Certified. Shall be heavy duty, waterproof, hospital grade waterproof plus. Completely enclosed motor can be totally immersed in water. Brand/Model No. **Salam** Electric Autopsy Saw, **Model No. 1200**, 115 Vac 60 Hz or approved equal.

Question 2: Can you please tell me how much you paid for Bone Saw last time? Do you have a bid tabulation from last time?

Answer: The price was \$3,132.00. See attached bid tabulation.

ADDENDUM RECEIVED AND ACKNOWLEDGED

Name of Firm

Representative

Signature

Date

QUOTE NO. 33513Q0001, AUTOPSY SAW

DESCRIPTION	BRAND	QTY	SOUTHLAND MEDICAL (STRYKER INSTR)		SALAM INT'L (SALAM)		MIDWEST CONST. SERV. (MORTECH)	
			UNIT	TOTAL	UNIT	TOTAL	UNIT	TOTAL
AUTOPSY SAW		3	\$ 1,965.00	\$ 5,895.00	\$ 1,044.00	\$ 3,132.00	\$ 2,185.00	\$ 6,555.00
FREIGHT		1	55.00	55.00	INCLUDED	-	INCLUDED	-
TOTAL			\$ 5,950.00		\$ 3,132.00		\$ 6,555.00	
DELIVERY			10 DAYS		5 DAYS		5 DAYS	